

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D1096634	(X3) Date Survey Completed 12/04/2024
Name of Provider or Supplier Ameripath Lubbock 501(A)	Street Address, City, State 2001 West Rosedale Street, Fort Worth, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was found to be in substantial compliance with CLIA regulations 42 CFR Part 493. Standard level deficiencies were cited.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observations, review of the CMS (Center for Medicare and Medicaid Services) 116 form, and confirmed in staff interview, the laboratory failed to ensure two of two tissue marking stains did not exceed their expiration dates. Findings included: 1. During a tour of the histopathology laboratory on 12/04/2024 at 10:54 a.m., the surveyor observed the following expired tissue marking stains located on the counter next to the microscope: Tissue Marking Dye - Green; lot #9253; Expiration date: 09/30/2021 Tissue Marking Dye - Black; lot #9254; Expiration date: 09/30/2021 2. Based on review of the CMS 116 form submitted on the date of the survey, the laboratory performed 17 histopathology frozen sections annually. 3. During an interview 12/04/2024 at 10:56 a.m. in the laboratory, the Quality Assurance Coordinator confirmed the tissue marking stains had exceeded their expiration date.</p>