

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D1099774	(X3) Date Survey Completed 10/08/2019
Name of Provider or Supplier Surgical Dermatology Associates	Street Address, City, State 4851 South I-35, Suite 101, Corinth, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was found to be in substantial compliance for the specialties /subspecialties for which it was surveyed. Exit conference was held with the laboratory manager on 10/08/2019. The standard level deficiency cited was discussed. The process for submitting the corrections was explained. CMS form 2567 will be emailed from the Texas Health and Human Services Commission, Patient Quality Arlington Group.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and in interview with staff, the laboratory failed to ensure histopathology reagents were not beyond their expiration date. Findings included: 1. During a tour of the laboratory on 10/08/2019 at 2:10 pm, the following reagents were observed to be stored in a cabinet: Vintage Bluing Reagent (1 Gallon) Lot #30167 expiration date 05/17 HistoPrep Xylenes (1 Gallon) Lot #146468 expiration date 10 /2012 (received 12/21/09) 2. During an interview on 10/08/2019 at 2:10 pm, the laboratory manager stated they are a green laboratory now and don't use those reagents. She reviewed and confirmed the above findings.</p>