

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D1099774	(X3) Date Survey Completed 05/28/2025
Name of Provider or Supplier Surgical Dermatology Associates	Street Address, City, State 4851 South I-35, Suite 101, Corinth, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was found to be in substantial compliance with CLIA regulations 42 CFR Part 493. Standard level deficiencies were cited.
D6128	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>(b)(9) Thereafter, evaluations must be performed at least annually unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individuals performance must be reevaluated to include the use of the new test methodology or instrumentation.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory policy, personnel records, and confirmed in staff interview, the technical supervisor failed to evaluate and document annual competency assessments for two of two testing persons (TP-1 and TP-2) responsible for high complexity testing in 2023 and 2024. Findings included: 1. Review of laboratory policy titled "JOB DESCRIPTION" stated: "JOB TITLE: Technical Supervisor DUTIES AND RESPONSIBILITIES: Listed but not limited to ... 4. Daily responsibilities that may be delegated to the Lab Manager ... d. Evaluate and document the performance of all testing personnel" 2. Review of personnel records revealed the technical supervisor failed to evaluate and document annual competency assessment for TP-1 and TP-2 in 2023 and 2024. 3. During an interview on 05/28 /2025 at 10:12 a.m., the Laboratory Manager, confirmed the technical supervisor failed to evaluate and document annual competency assessments for two of two testing persons (TP-1 and TP-2) responsible for high complexity testing in 2023 and 2024.</p>