

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D1103581	(X3) Date Survey Completed 07/24/2025
Name of Provider or Supplier Westlake Dermatology & Cosmetic Surg	Street Address, City, State 800 S Highway 281, Marble Falls, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was surveyed and found to be in compliance with the Conditions of the CLIA regulations found at 42 CFR 493.1 through 493.1780, and recertification is recommended. Standard level deficiency was cited.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory policy and procedure, accuracy assessments, pre-survey paperwork, and interview, the laboratory failed to perform twice per year accuracy assessments for Mohs in 2023 for one of three events reviewed. Findings follow. A. Review of the laboratory's policy and procedure titled Comparison of Test Results Policy, under Procedure stated, "Twice a year, any test that is not covered by an external proficiency test must be checked for accuracy and reliability." B. Review of accuracy assessments from 2023 and 2024 showed one accuracy assessment in 2023. Additional accuracy assessments were requested on July 24, 2025 at 1445 hours but not provided. C. Review of the pre-survey paperwork showed approximately 536 Mohs blocks were performed per year. D. Interview with the histotechnologist on July 24, 2025 at 1445 hours in the office confirmed they were missing an accuracy assessment in 2023</p>