

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2000387	(X3) Date Survey Completed 02/08/2018
Name of Provider or Supplier Texas Internal Medicine And Diagnostic Center	Street Address, City, State 4502 N Laurent St, Victoria, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	<p>A follow-up visit for the proficiency test desk review survey was performed onsite on February 8, 2018. The laboratory failed to cease testing for the analytes of Hematocrit and Red Blood Cell according to the enforcement letter dated July 5, 2017. The laboratory performed 319 CBC (complete blood count) patient tests that included the analytes of Hematocrit and Red Blood Cell from the period of July 20, 2017 to February 8, 2018. The laboratory has not demonstrated two consecutive sustained satisfactory performance (>80%) on two consecutive testing events for reinstatement. The laboratory continued to be out of compliance based on the following CONDITION LEVEL DEFICIENCIES: 493.803 successful participation in a proficiency testing program 493.807 (a) Reinstatement after Failure 493.1403 laboratories performing moderate complexity testing; laboratory director ***** 30468 The following deficiencies are a result of two desk reviews performed on October 3, 2017 and November 16, 2017 of proficiency testing scores obtained from the CMS (Center for Medicare Services) national database and verified with the proficiency testing company, American Proficiency Institute (API). The facility was found to be out of compliance with the conditions of participation of the CLIA program. The following CONDITION LEVEL DEFICIENCIES were found to be out of compliance: 493.803 successful participation in a proficiency testing program 493.1403 laboratories performing moderate complexity testing; laboratory director 493.807 (a) reinstatement after failure These are repeat deficiencies noted on the 6/17/17 CMS 2567.</p>