

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D2015888	<b>(X3) Date Survey Completed</b>  04/10/2018
<b>Name of Provider or Supplier</b>  Angiocardiac Care Of Tx	<b>Street Address, City, State</b>  10019 Main St A-1, Houston, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	<p>Noted deficiencies and plans of correction were discussed with the laboratory representative(s) at the exit conference. The facility representative(s) were given an opportunity to provide evidence of compliance with the noted deficiencies, and no such evidence was provided prior to survey exit. The facility was found to be in compliance with applicable Conditions of Participation in the CLIA program, and certification is recommended. Note: The CMS-2567 (Statement of Deficiencies) is an official, legal document. All information must remain unchanged except for entering the plan of correction, correction dates, and the signature space. Any discrepancy in the original deficiency citation(s) will be reported to the Dallas Regional Office (RO) for referral to the Office of the Inspector General (OIG) for possible fraud. If information is inadvertently changed by the provider/supplier, the State Survey Agency (SA) should be notified immediately.</p>
<b>D1001</b>	<p><b>CERTIFICATE OF WAIVER TESTS</b> CFR(s): 493.15(e)</p> <p>Laboratories eligible for a certificate of waiver must-- (1) Follow manufacturers' instructions for performing the test; and (2) Meet the requirements in subpart B, Certificate of Waiver, of this part.</p> <p>This STANDARD is not met as evidenced by: Based on observation, review of package inserts and interview of laboratory personnel it was revealed that the laboratory failed to document the room temperature of the medical record room where urine drug testing is performed. Findings were: 1. During a tour of the facility, it was observed that there was no chart for recording room temperature in the medical record room where urine drug testing is performed. 2. Observation of laboratory supplies stored at room temperature revealed the following supplies: DiscoverPlus Multipanel Drug Test (2-30 C). 3. An interview of the medical assistant on 04/10/2018 at 0935 hours in medical record room confirmed the above findings.</p>