

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2030884	(X3) Date Survey Completed 01/18/2022
Name of Provider or Supplier Epiphany Dermatology, Pa	Street Address, City, State 1038 S Fleishel Ave, Tyler, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	. An onsite survey conducted on 1/18/2022 found the laboratory in compliance with 42 CFR Part 493, Requirements for Laboratories. .
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: . Based on review of laboratory final reports and confirmed in interview, the laboratory failed to include the address of the laboratory location where the test was performed on six of six Mohs maps, included in the patient final report, reviewed in 2020 and 2021. The findings include: 1. Review the following six patient final reports from 2020 and 2021 included a 'Mohs Micrographic Map' worksheet that did not include the address of the laboratory location where the testing was performed; 2020 - Three patient final reports reviewed: M20-566 M20-580 M20-535 2021 - Three patient final reports reviewed: M21-628 M21-636 M21-601 2. In an interview on 1/18 /2022 at 11:30 hours, testing personnel (TP) 4 confirmed the above findings. .</p>