

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2047878	(X3) Date Survey Completed 03/11/2021
Name of Provider or Supplier Ralph Alhalel Pa	Street Address, City, State 1200 E Ridge Road, Suite 5, Mcallen, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	Noted deficiencies and plans of correction were discussed with the laboratory representative at the entrance and exit conferences. The facility representative was given an opportunity to provide evidence of compliance with the noted deficiencies, and no such evidence was provided prior to survey exit. The facility was found to be in compliance with applicable Conditions of Participation in the CLIA program, and recertification is recommended. Note: The CMS-2567 (Statement of Deficiencies) is an official, legal document. All information must remain unchanged except for entering the plan of correction, correction dates, and the signature space. Any discrepancy in the original deficiency citation(s) will be reported to the Dallas Regional Office (RO) for referral to the Office of the Inspector General (OIG) for possible fraud. If information is inadvertently changed by the provider/supplier, the State Survey Agency (SA) should be notified immediately.
D6143	GENERAL SUPERVISOR QUALIFICATIONS CFR(s): 493.1461 (a) The general supervisor must possess a current license issued by the State in which the laboratory is located, if such licensing is required; and (b) The general supervisor must be qualified as a-- (b)(1) Laboratory director under 493.1443; or (b)(2) Technical supervisor under 493.1449. (c) If the requirements of paragraph (b)(1) or paragraph (b)(2) of this section are not met, the individual functioning as the general supervisor must-- (c)(1)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located or have earned a doctoral, master's, or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; and (c)(1)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing; or (c)(2)(i) Qualify as testing personnel under 493.1489(b)(2); and (c)(2)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing; or (c)(3)(i) Except as specified in paragraph (3)(ii) of this section, have previously qualified as a general supervisor under 493.1462 on or before February 28, 1992. (c)(3)(ii)

Exception. An individual who achieved a satisfactory grade in a proficiency examination for technologist given by HHS between March 1, 1986 and December 31, 1987, qualifies as a general supervisor if he or she meets the requirements of 493.1462 on or before January 1, 1994. (c)(4) On or before September 1, 1992, have served as a general supervisor of high complexity testing and as of April 24, 1995-- (c)(4)(i) Meet one of the following requirements: (c)(4)(i)(A) Have graduated from a medical laboratory or clinical laboratory training program approved or accredited by the Accrediting Bureau of Health Education Schools (ABHES), the Commission on Allied Health Education Accreditation (CAHEA), or other organization approved by HHS. (c)(4)(i)(B) Be a high school graduate or equivalent and have successfully completed an official U.S. military medical laboratory procedures course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician). (c)(4)(ii) Have at least 2 years of clinical laboratory training, or experience, or both, in high complexity testing; or (c)(5) On or before September 1, 1992, have served as a general supervisor of high complexity testing and-- (c)(5)(i) Be a high school graduate or equivalent; and (c)(5)(ii) Have had at least 10 years of laboratory training or experience, or both, in high complexity testing, including at least 6 years of supervisory experience between September 1, 1982 and September 1, 1992. (d) For blood gas analysis, the individual providing general supervision must-- (d)(1) Be qualified under 493.1461(b)(1) or (2), or 493.1461(c); or (d)(2)(i) Have earned a bachelor's degree in respiratory therapy or cardiovascular technology from an accredited institution; and (d)(2)(ii) Have at least one year of laboratory training or experience, or both, in blood gas analysis; or (d)(3)(i) Have earned an associate degree related to pulmonary function from an accredited institution; and (d)(3)(ii) Have at least two years of training or experience, or both in blood gas analysis. (e) The general supervisor requirement is met in histopathology, oral pathology, dermatopathology, and ophthalmic pathology because all tests and examinations, must be performed: (e)(1) In histopathology, by an individual who is qualified as a technical supervisor under 493.1449(b) or 493.1449(l)(1); (e)(2) In dermatopathology, by an individual who is qualified as a technical supervisor under 493.1449(b) or 493.1449(l) or (2); (e)(3) In ophthalmic pathology, by an individual who is qualified as a technical supervisor under 493.1449(b) or 493.1449(1)(3); and (e)(4) In oral pathology, by an individual who is qualified as a technical supervisor under 493.1449(b) or 493.1449(m).

This STANDARD is not met as evidenced by:

Based on review of the laboratory's records, and staff interview, it was revealed the laboratory failed to have documentation of the general supervisor's review of grossing. The findings were: 1. A review of the laboratory's records revealed the laboratory failed to have documentation of a policy stating the general supervisor would review tissues and blocks within 24 hours of grossing. 2. The laboratory was asked to provide documentation of the general supervisor reviewing tissue and blocks within 24 hours of grossing being performed. No documentation was provided. 3. An interview with the Chief Operations Officer on March 11, 2021 at 11:45 hours in the break room confirmed the findings.