

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2056204	(X3) Date Survey Completed 10/12/2023
Name of Provider or Supplier Crescent Medical Center Lancaster	Street Address, City, State 2600 West Pleasant Run Road, Lancaster, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the CMS (Centers for Medicare and Medicaid Services) national database and verified with the proficiency testing company, American Proficiency Institute (API). The facility was found to be out of compliance with CLIA regulations 42 CFR Part 493. CONDITION LEVEL DEFICIENCIES were found to be out of compliance: 493.803 Successful participation [proficiency testing] 493.1441 Laboratories performing high complexity testing; laboratory director
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing records, the laboratory failed to</p>

	<p>successfully participate in a proficiency testing program approved by HHS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory failed to successfully participate in the specialty of Immunohematology for Compatibility Testing. Refer to D2181.</p>
<p>D2181</p>	<p>COMPATIBILITY TESTING CFR(s): 493.863(e)</p> <p>Failure to achieve an overall testing event score of satisfactory for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of the CASPER- 0155 Individual Laboratory Report obtained from the CMS national database and API 2022 (3rd event) and 2023 (2nd event) records, the laboratory failed to achieve satisfactory performance (100%) for the same analyte in two out of three consecutive testing events in the specialty of Immunohematology for Compatibility Testing. Two of three consecutive unsatisfactory scores result in unsuccessful PT performance. The findings include: 1. Review of the CASPER- 0155 report revealed the following: Compatibility Testing 2022 - 3rd Event Laboratory received an unsatisfactory score of 60% Compatibility Testing 2023 - 2nd Event Laboratory received an unsatisfactory score of 80% 2. A proficiency testing desk review from API 2022 and 2023 proficiency testing records confirmed the above findings. Key: CASPER: Certification and Survey Provider Enhanced Reporting CMS: Centers for Medicare and Medicaid API: American Proficiency Institute</p>
<p>D6076</p>	<p>LABORATORY DIRECTOR CFR(s): 493.1441</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of laboratory proficiency testing performance it was revealed that the laboratory director failed to provide overall management and direction of the laboratory services. Refer to D6089.</p>
<p>D6089</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(i)</p> <p>The laboratory director must ensure the proficiency testing samples are tested as required under subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a desk review of proficiency testing results it was revealed that the laboratory director failed to ensure the overall quality of the laboratory services provided. The laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2181.</p>