

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 45D2059871	<b>(X3) Date Survey Completed</b> 05/19/2026
<b>Name of Provider or Supplier</b> Ari Brown Pediatrics,Pa	<b>Street Address, City, State</b> 925 Westbank Drive, Suite 100, Austin, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The Ari Brown Pediatrics DBA 411 Pediatrics laboratory was found to be in compliance with the Conditions of the CLIA regulations found at 42 CFR 493.1 through 493.1780, CLIA requirements for laboratories as a result of a recertification survey on 05/19/2026 and recertification is recommended. Standard level deficiency was cited.
<b>D6053</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(9)</p> <p>(b)(9) Evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's duties and responsibilities, training records, competency evaluations, and interview, the technical consultant failed to evaluate the competency at least semiannually during the first year the individual tested patient specimens for one of four new employees for the Complete Blood Count (CBC) using the Medonic M Series analyzer. Findings follow. A. Review of the duties and responsibilities for the Technical Consultant, 07/22/2013, stated, "10. Evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens..." There was no policy and procedure that covered competency evaluations. B. Review of the Testing Personnel Training Record for testing personnel #3 (as listed on the CMS form 209) started patient testing 02/21/2024. C. Review of the Personnel Assessment showed competencies were performed 07/21/2024 and 08/05/2025 (elapsed time 1 year 5 months 15 days). A second semi-annual competency evaluation for testing personnel #3 was requested on May 19, 2026 at 1240 hours but not provided. D. Interview with the Technical Consultant on May 19, 2026 at 1240 hours confirmed the findings.</p>