

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2059958	(X3) Date Survey Completed 11/09/2018
Name of Provider or Supplier Kids Kare Pediatrics	Street Address, City, State 5019 Portico Way, Midland, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on a review of laboratory policies and procedures and interview of facility personnel it was revealed that the laboratory failed to have a procedure to assess the competency of all consultants and testing personnel . Findings were: 1. Review of the laboratory policies and procedures found a policy/ procedure titled Competency Assessment of Testing Personnel. There was not a policy/procedure for evaluating the competency of the technical consultant. A policy for assessing the competency of the Technical consultant was requested but not provided. 2. Interview of Testing person one conducted on November 9, 2018 at 09:06 AM confirmed there were no competency assessment records or a procedure for assessing the competency of technical consultants .</p>