

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D2064081	<b>(X3) Date Survey Completed</b>  07/18/2019
<b>Name of Provider or Supplier</b>  Complete Care Camp Bowie	<b>Street Address, City, State</b>  6006 Camp Bowie, Fort Worth, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The Complete Care Camp Bowie clinical laboratory is not in compliance with the 42 CFR Part 493, Requirements for Laboratories. Biennial certification survey was conducted 07/18/19. Standard level deficiencies were cited as follows:
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of American Proficiency Institute (API) proficiency testing and interview with the Technical Consultant the laboratory failed to score at least 80% for White Blood Cell (WBC) differential for 1 (3rd testing event 2017) out of 6 (1st, 2nd testing event 2019, 1st, 2nd, 3rd testing event 2018, and 3rd testing event 2017) testing events reviewed. Findings Included: Review of API proficiency testing for WBC differential revealed a score of 0% in the 3rd testing event in 2017. Interview on 07/18/19 at 10:30 AM the Technical Consultant confirmed the proficiency testing failure in the 3rd testing event of 2017 in WBC differential.</p>