

| | | |
|---|---|--|
| <p>Statement of Deficiencies</p> | <p>(X1) Provider/Supplier/CLIA Identification Number 45D2082207</p> | <p>(X3) Date Survey Completed 11/05/2018</p> |
| <p>Name of Provider or Supplier Mcguiness Dermatology Center Of Flower Mound</p> | <p>Street Address, City, State 4471 Long Prairie Road Suite 100, Flower Mound, TX</p> | |
| <p>For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.</p> | | |

| <p>(X4) ID Prefix Tag</p> | <p>Summary Statement of Deficiencies</p> |
|----------------------------------|--|
| <p>D5209</p> | <p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Office Manager the laboratory failed to perform competency assessments on 4 out of 4 Testing Personnel who performs KOH (Potassium Hydroxide) and Scabies testing. Findings Included: Review of the CMS 209 revealed 4 Testing Personnel who performed the KOH and Scabies testing. Record review revealed no documentation of competency assessments for all 4 Testing Personnel. During an interview on 11/05/18 at 3:08 PM the Office Manager confirmed that no competency assessments had been performed on the 4 Testing Personnel.</p> |
| <p>D5217</p> | <p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Office Manager the laboratory failed to verify the accuracy of the KOH (Potassium Hydroxide) and Scabies testing for 2 out of 2 years (2016-2017) reviewed. Findings Included: Review of peer review records revealed no peer review was performed for KOH or Scabies testing from 10</p> |

/2016 to 10/2018. During an interview on 11/05/18 at 1:43 PM the Office Manager confirmed that there was no peer review documented for the KOH or Scabies testing.