

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D2087522	<b>(X3) Date Survey Completed</b>  08/02/2024
<b>Name of Provider or Supplier</b>  Pioneer Lab Houston Lp	<b>Street Address, City, State</b>  9130 South Texas 6, Houston, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The laboratory was surveyed and found to be in compliance with the Conditions of the CLIA regulations found at 42 CFR 493.1 through 493.1780, and recertification is recommended. Standard level deficiencies were cited.
<b>D5805</b>	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's records, a random review of patient's test reports from July 2024, and staff interview, the laboratory failed to include the correct address of the testing facility on three of three patient's test reports reviewed. Findings include: 1. A review of the laboratory's records revealed the laboratory moved in April 2024. 2. A random review of patient's test reports from July 2024 revealed the laboratory failed to include the correct address of the testing facility on the following 3 patient's initial screening test reports: Patient ID: A03151 Test Date: 7/31/24 Patient ID: A03095 Test Date: 7/31/24 Patient ID: A03152 Test Date: 7/31/24 3. Further review of the patient's initial screening test reports revealed the reports listed the previous address instead of the current address for the laboratory. 4. In an interview on 8/2/24 at 11:25 a.m. in the laboratory, after review of the records, the compliance consultant confirmed the above findings.</p>