

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2095666	(X3) Date Survey Completed 11/05/2020
Name of Provider or Supplier Us Dermatology Partners	Street Address, City, State 300 Richland West Circle, Suite 2c, Waco, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's quality assurance records from 2018 through 2020 and staff interview, it was revealed the laboratory failed to have documentation of performing twice annual accuracy assessment for KOH preparations in 2018 and 2019. Findings included: 1. A review of the laboratory's Parasitology and KOH Competency Exams from 2018 through 2020 found documentation of verification of the accuracy of results for KOH preparations done once each year in 2018 and 2019. The laboratory was not enrolled in a CMS approved proficiency testing program for Clinical Microscopy. 2. An interview with the clinic manager conducted on November 5, 2020 at 10:45 a.m. confirmed the laboratory was neither enrolled in a proficiency testing program nor did the laboratory verify the accuracy of the test results for KOH at least twice a year in 2018 or 2019. Key: KOH - Postassium hydroxide CMS - Centers for Medicare and Medicaid Services</p>