

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2104701	(X3) Date Survey Completed 01/07/2026
Name of Provider or Supplier Complete Emergency Care La Vernia Llc	Street Address, City, State 102 S Fm 1346, Suite 2, La Vernia, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the proficiency testing company. The facility was found to be out of compliance with the conditions of the CLIA program. The following CONDITION LEVEL DEFICIENCIES were found: D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing] D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; laboratory director
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting

	(CASPER) Report 155 Individual Laboratory Profile and American Proficiency Institute's proficiency reports, the laboratory failed to achieve satisfactory performance in two of two testing events for the analyte of CK, isoenzymes (refer to D2096) resulting in an initial unsuccessful performance.
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f) (f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: 1. A review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile report, the laboratory received the following unsatisfactory performances for CK, isoenzymes on two of two events: 2025 Event 2 40% 2025 Event 3 60% 2. A review of the American Proficiency Institute's proficiency testing reports from 2025 confirmed the findings.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on proficiency testing desk review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and the American Proficiency Institute's proficiency testing reports from 2025 (Events 2 and 3), the laboratory director failed to provide overall management and direction of the laboratory services resulting in an initial proficiency testing failure for the analyte CK, isoenzymes (refer to D6016.)
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) (e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on proficiency desk review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and the American Proficiency Institute's proficiency testing reports from 2025 (Events 2 and 3), the laboratory director failed to ensure successful participation in a HHS approved proficiency testing program for the analyte of CK, isoenzymes for two of two events in 2025, resulting in an initial unsuccessful performance (refer to D2096).