

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2114025	(X3) Date Survey Completed 09/03/2019
Name of Provider or Supplier Octapharma Plasma Inc	Street Address, City, State 11799 West Avenue, San Antonio, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	Noted deficiencies and plans of correction were discussed with the laboratory representative at the entrance and exit conferences. The facility representative was given an opportunity to provide evidence of compliance with the noted deficiencies, and no such evidence was provided prior to survey exit. The facility was found to be in compliance with applicable Conditions of Participation in the CLIA program, and recertification is recommended. Note: The CMS-2567 (Statement of Deficiencies) is an official, legal document. All information must remain unchanged except for entering the plan of correction, correction dates, and the signature space. Any discrepancy in the original deficiency citation(s) will be reported to the Dallas Regional Office (RO) for referral to the Office of the Inspector General (OIG) for possible fraud. If information is inadvertently changed by the provider/supplier, the State Survey Agency (SA) should be notified immediately.
D5441	<p>CONTROL PROCEDURES CFR(s): 493.1256(a)(b)(c)(g)</p> <p>(a) For each test system, the laboratory is responsible for having control procedures that monitor the accuracy and precision of the complete analytic process. (b) The laboratory must establish the number, type, and frequency of testing control materials using, if applicable, the performance specifications verified or established by the laboratory as specified in 493.1253(b)(3). (c) The control procedures must-- (c)(1) Detect immediate errors that occur due to test system failure, adverse environmental conditions, and operator performance. (c)(2) Monitor over time the accuracy and precision of test performance that may be influenced by changes in test system performance and environmental conditions, and variance in operator performance. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's records, and staff interview, it was revealed the laboratory failed to have documentation of monitoring quality control values over</p>

time for accuracy and precision of test performance. The findings were: 1. A review of the laboratory's quality control records from 2018 and 2019 revealed the laboratory tested two levels of control material each day of patient testing for total protein on each of its eight refractometers. 2. The laboratory was asked to provide documentation of monitoring the quality control values over time. No documentation was provided. 3. An interview with the Quality Assurance Supervisor on 09/03/2019 at 1100 hours in the training room revealed the laboratory monitored quality control values daily for acceptability, but did not monitor the values over time for accuracy and precision. This confirmed the findings.