

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2116798	(X3) Date Survey Completed 07/23/2024
Name of Provider or Supplier Dermisurgery Associates-The Woodlands	Street Address, City, State 17183 I-45 Suite 510, Conroe, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced survey of the laboratory was conducted on 07/23/2024. The laboratory was found in compliance with applicable CLIA regulations (42 CFR Part 493, Requirements for Laboratories) for the specialties/subspecialties for which it was surveyed. STANDARD LEVEL DEFICIENCIES were cited.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's observations, review of laboratory's annual test volumes and staff interview, the laboratory failed to ensure 4 of 4 marking dyes used for Mohs procedures were not used after exceeding their expiration date. Findings included: 1. Surveyor's observations on 07/23/2024 at 1310 hours in the laboratory revealed the following marking dyes used for laboratory procedures exceeded their expiration date: Chemtrec Blue Tissue Marking Dye Lot: 148208 Expired: 2024-04-30 Chemtrec Orange Tissue Marking Dye Lot: 148901 Expired: 2024-04-30 Cancer Diagnostics Inc. Blue Tissue Marking Dye Lot: 21133 Expired: 2023-05-31 Cancer Diagnostics Inc. Black Tissue Marking Dye Lot: 21152 Expired: 2023-06-30 2. Review of laboratory's submitted annual test volumes revealed the laboratory performed 1000 Mohs procedures annually. 3. In an interview on 07/23/2024 at 1320 hours in the laboratory, the laboratory's Histotechnologist on duty, after review of the expired products, confirmed the findings.</p>