

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2118967	(X3) Date Survey Completed 10/23/2018
Name of Provider or Supplier Texas Surgical Dermatology Pa	Street Address, City, State 21009 Kuykendahl Road Suite A, Spring, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory records from 2017 through 2018 and staff interview, it was revealed the laboratory failed to have documentation of performing twice annual accuracy assessment for MOHS testing. Findings included: 1. A review of the laboratory's records from 2017 through 2018 found the laboratory had no documentation for performing twice annual accuracy assessment of MOHS testing available for review. 3. Interview of the laboratory director conducted on October 23,2018 at 1:49 PM confirmed that there was no documentation available for review of semi-annual accuracy assessment of MOHS testing. He stated that he had participated in the 2017 MOHS College Peer review in 2017 but did not have that available for review, and did not send out slides for peer review.</p>