

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2148519	(X3) Date Survey Completed 03/27/2023
Name of Provider or Supplier Sanova Dermatology-Bee Cave	Street Address, City, State 3944 S Fm 620 S, Bldg 6, Suite 201, Bee Cave, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was surveyed and found to be in compliance with the Conditions of the CLIA regulations found at 42 CFR 493.1 through 493.1780, and recertification is recommended.
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on review of the patient testing log, patient test report, and interview, the laboratory failed to ensure the correct result was reported for KOH (potassium hydroxide) test for fungal elements for 1 of 3 test reports reviewed from Oct 2022 - March 2023. Findings follow. Review of the patient testing log from 10/12/22 - 03/17/23 against the patient test report showed 1 of 3 with the incorrect result. Review of the KOH log showed patient tested on 03/17/2023, MRN TK4GOytgNr2B, was reported as negative. Review of the patient's chart on 03/17/2023 showed the KOH was reported as branching hyphae. Interview with the Mohs technologist by phone on March 24, 2023, at 1610 hours confirmed the result was negative and incorrectly reported as positive in the chart by a medical assistant.</p>