

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2149371	(X3) Date Survey Completed 05/11/2022
Name of Provider or Supplier Crockett Medical Center	Street Address, City, State 1100 East Loop 304, Crockett, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	Based on a proficiency testing desk review survey performed on May 11, 2022, the laboratory was found to be out of compliance based on the following CONDITION LEVEL DEFICIENCIES : D2016 - 42 C.F.R. 493.803 Condition: Successful participation D6000 - 42 C.F.R. 493.1403 Condition: Laboratory Director, moderate complexity
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile, and American Proficiency Institute (API) proficiency testing records from 2021 and 2022, the laboratory failed</p>

	<p>to achieve satisfactory performance (80% or greater) for the analyte Partial Thromboplastin Time (APTT) in three out of three consecutive events, resulting in unsuccessful performance. Refer to D2130.</p>
<p>D2130</p>	<p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on desk review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile, and American Proficiency Institute (API) proficiency testing records from 2021 and 2022, the laboratory failed to achieve satisfactory performance (80% or greater) for the analyte Partial Thromboplastin Time (APTT) in three out of three consecutive events. The findings include: 1. A review of the CASPER report 155 lists a score of "40%" for the API Hematology 2nd Event, a score of "60%" for the API Hematology 3rd Event, and a score of 60% for the 1st event of 2022 for the analyte APTT. 2. A proficiency desk review of the API proficiency testing records from 2021 confirmed that the laboratory received unsatisfactory scores for APTT during the API Hematology 2nd and 3rd Event of 2021 and 1st event of 2022.</p>
<p>D6000</p>	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on desk review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and American Proficiency Institute (API) proficiency testing records, the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program for analyte APTT. Refer to D6016.</p>
<p>D6016</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on desk review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and American Proficiency</p>

Institute (API) proficiency testing records, the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program for analyte APTT for three out of three events in 2021 and 2022. Refer to D2130.