

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2172775	(X3) Date Survey Completed 02/20/2025
Name of Provider or Supplier Elite Dermatology	Street Address, City, State 20326 Tomball Parkway, Suite #400, Houston, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced survey of the laboratory was conducted on 02/20/2025. The laboratory was found in compliance with applicable CLIA regulations (42 CFR Part 493, Requirements for Laboratories) for the specialties/subspecialties for which it was surveyed. STANDARD LEVEL DEFICIENCIES were cited.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's observations, review of laboratory's annual test volumes and staff interview, the laboratory failed to ensure four of five marking dyes used for Mohs procedures were not used after exceeding their expiration date. Findings included: 1. Surveyor's observations on 02/20/2025 at 0905 hours in the laboratory revealed the following marking dyes used for laboratory procedures exceeded their expiration date: Avantik Green Tissue Marking Dye Lot: 151409 Expired: 2024-11-30 Avantik Blue Tissue Marking Dye Lot: 156628 Expired: 2024-08-31 Avantik Red Tissue Marking Dye Lot: 159204 Expired: 2024-10-31 Avantik Black Tissue Marking Dye Lot: 157822 Expired: 2024-09-30 2. Review of laboratory's submitted annual test volumes revealed the laboratory performed approximately 500 test procedures annually. 3. In an interview on 02/20/2025 at 0910 hours in the laboratory, the laboratory's Chief Operating Officer (as indicated on submitted Entrance/Exit Conference document) confirmed the findings.</p>