

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2196860	(X3) Date Survey Completed 03/23/2022
Name of Provider or Supplier Realskin Dermatology	Street Address, City, State 6612 Woodway Dr., Building A, Suite 2, Waco, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's policies and procedures, testing logs, and interview, the laboratory failed to verify the accuracy of its KOH (potassium hydroxide) for fungal elements and Mineral Oil for Scabies tests at least twice annually. Findings follow. A. Review of the laboratory's policy and procedure General Laboratory Quality Systems under Proficiency Testing Policy stated, "For Provider Performed Microscopies, the laboratory will either be enrolled in proficiency testing or will perform split-specimen testing with another laboratory between the providers. The provider's results will be evaluated by the laboratory director. If they compare well, they will be considered 'passing.' If the provider's results do not agree well, it will be considered 'unsuccessful' and will be treated accordingly." B. 1. Review of the Provider-Performed Microscopy Log showed testing personnel #3 first tested KOH for fungal elements on 01/11/2021 and testing personnel #4 first tested KOH for fungal elements on 01/26/2021. 2. Review of the Provider-Performed Microscopy Log showed testing personnel #3 first tested for Scabies on 05/03/2021 and testing personnel #4 first tested for Scabies on 04/30/2021. C. Interview with the practice manager on March 22, 2022 at 0955 in the office confirmed they did not have documentation of twice per year accuracy assessments for KOH and Scabies.</p>