

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2217639	(X3) Date Survey Completed 07/15/2024
Name of Provider or Supplier Gastroenterology Consultants Of Laredo Pathology	Street Address, City, State 6999 Mcperson Road, Suite 211-C, Laredo, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's accuracy assessments from 2022 and 2023 for the professional component for H&E, HP Blue and ABPAS stains, and staff interview, the laboratory failed to have documentation of performing the assessments 1 of 2 times in 2022 and 1 of 2 times in 2023. The findings included: 1. A review of the laboratory's accuracy assessments from 2022 and 2023 determined assessments were performed at the following times: June 2022 June 2023 2. The laboratory did not have documentation of performing the assessments twice annually in 2022 and 2023. 3. The histotech acknowledged the facility only performed accuracy assessments once per year in 2022 and 2023. She was stated she was not aware the accuracy assessments were to be performed twice annually. This confirmed the findings.</p>