

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 45D2218596	<b>(X3) Date Survey Completed</b> 12/04/2025
<b>Name of Provider or Supplier</b> Longhorn Dermatology	<b>Street Address, City, State</b> 4900 Bee Creek Rd #101, Spicewood, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The Longhorn Dermatology laboratory was found to be in compliance with the Conditions of the CLIA regulations found at 42 CFR 493.1 through 493.1780, CLIA requirements for laboratories as a result of a recertification survey on December 4, 20205 and recertification is recommended. Standard level deficiencies were cited.
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation, manufacturer's instructions, pre-survey paperwork, and interview, the laboratory failed to ensure reagents were not used beyond their expiration for four of seven tissue marking dyes currently in use by the laboratory used in Mohs testing. Findings follow. A. Surveyor observed on December 4, 2025 at 1500 hours in the laboratory four of seven Cancer Diagnostic Tissue Marking Dyes available and in use by the laboratory were expired as listed by color, lot number, and expiration date: Color Lot # Expiration date 1. Orange 23082 03/31/2025 2. Red 23122 05/31/2025 3. Purple 23180 06/30/2025 4. Yellow 23258 09/30/2025 B. Interview with the histotech on December 4, 2025 at 1505 hours in the laboratory confirmed the findings. C. Review of the CMS form 116 showed approximately 2011 blocks were performed annually.</p>