

<p>Statement of Deficiencies</p>	<p>(X1) Provider/Supplier/CLIA Identification Number</p> <p>45D2218613</p>	<p>(X3) Date Survey Completed</p> <p>08/19/2024</p>
<p>Name of Provider or Supplier</p> <p>Radiant Dermatology & Aesthetics PLLC</p>	<p>Street Address, City, State</p> <p>9240 N Sam Houston Pkwy East, Suite 201, Humble, TX</p>	
<p>For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.</p>		

<p>(X4) ID Prefix Tag</p>	<p>Summary Statement of Deficiencies</p>
<p>D0000</p>	<p>An announced survey of the laboratory was conducted on 08/19/2024. The laboratory was found in compliance with applicable CLIA regulations (42 CFR Part 493, Requirements for Laboratories) for the specialties/subspecialties for which it was surveyed. STANDARD LEVEL DEFICIENCIES were cited.</p>
<p>D5417</p>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor's observations, review of laboratory's annual test volumes and staff interview, the laboratory failed to ensure 3 of 4 marking dyes used for Mohs procedures were not used after exceeding their expiration date. Findings included: 1. Surveyor's observations on 08/19/2024 at 0910 hours in the laboratory revealed the following marking dyes used for laboratory procedures exceeded their expiration date: Avantik Yellow Tissue Marking Dye Lot: 148630 Expired: 2024-04-30 Avantik Green Tissue Marking Dye Lot: 154057 Expired: 2024-07-31 Avantik Red Tissue Marking Dye Lot: 151000 Expired: 2024-06-30 2. Review of laboratory's submitted annual test volumes revealed the laboratory performed 674 Mohs procedures annually. 3. In an interview on 08/19/2024 at 0910 hours in the laboratory, the laboratory's Histotechnologist on duty, after review of the expired products, confirmed the findings.</p>