

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D2281375	(X3) Date Survey Completed 09/10/2025
Name of Provider or Supplier Define Dermatolgoy, Pllc	Street Address, City, State 4327 Barnett Rd, Wichita Falls, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was found to be in substantial compliance with CLIA regulations 42 CFR Part 493. Standard level deficiencies were cited.
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>(c) The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on review of patient Mohs maps and confirmed in interview, the laboratory failed to include the testing facility address on the patient Moh's maps for 10 of 10 patients in 2024 (random review December) and 12 of 12 patients in 2025 (random review September). Findings included: 1. A random review of patient Mohs maps from December 2024 and September 2025 revealed the following 22 patient Mohs maps which did not include the testing facility address: 12/5/2024 Patient Date of Birth: 09/10/1939, 08/21/1944, 05/18/1963, 01/11/1944, 10/12/1933, 03/11/1951, 08/07/1951, 05/26/1940, 06/11/1957, 09/16/1934 09/03/2025 Patient Date of Birth: 06/17/1949, 01/25/1952, 08/07/1949, 01/10/1939, 10/02/1940, 07/27/1957, 06/18/1954, 04/28/1963, 02/22/1957, 07/06/1958, 01/15/1947, 08/04/1946 2. During an interview on 09/10/2025 at 2:08 p.m., the Mohs histotechnician, after a review of records, confirmed the laboratory failed to include the testing facility address on the patient Moh's maps.</p>