

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0523765	(X3) Date Survey Completed 09/22/2020
Name of Provider or Supplier Wasatch Women's Center	Street Address, City, State 715 E 3900 S Ste 203, Salt Lake City, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least annually, after the first year.</p> <p>This STANDARD is not met as evidenced by: Based on competency records review, lack of documentation, and interview with staff, the laboratory technical consultant failed to evaluate the Rh type testing competency for 2 of 5 testing personnel for 1 of 2 years of testing performed (2019). Findings include: 1. Competency evaluation record review failed to include documentation Testing personnel C and D were evaluated in 2019 for Rh type competency through direct observation of Rh typing procedures, monitoring the recording and reporting of test results, review of intermediate worksheets and preventive maintenance records, assessment of test performance through proficiency testing samples, and assessment of problem solving skills. 2. In an interview with staff on 09/22/2020 at approximately 10:15 A.M. staff confirmed two staff members lacked 2019 competency evaluations.</p>