

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0525866	(X3) Date Survey Completed 10/28/2019
Name of Provider or Supplier Southwest Skin And Cancer	Street Address, City, State 383 S 300 E, St George, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with the director, the laboratory failed to ensure Potassium Hydroxide (KOH) reagent was not used past its expiration date of 04/13/2018. The laboratory performed approximately 3 to 5 KOH tests per month. Findings include: 1. The surveyor observed, on 10/28/2019 at approximately 7:40 A.M., one dropper bottle of KOH reagent, lot number 6104 with an expiration date of 04/13/2018 present at the microscope in the laboratory microscopic review area. 2. In an interview conducted on 10/29/2019 at approximately 11:45 A.M. the director confirmed the KOH reagent was in use past the expiration date.</p>