

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0525866	(X3) Date Survey Completed 09/25/2024
Name of Provider or Supplier Southwest Skin And Cancer	Street Address, City, State 383 S 300 E, St George, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with the lab manager, the laboratory failed to ensure that tissue marking dyes were not used past their expiration date. The laboratory performed approximately 13,000 histopathology tests annually. Findings Include: 1. Direct observation of tissue marking dyes on 09/25/2024 at 2:45 PM revealed that 3 of 3 tissue marking dyes had expired. The black TMD tissue marking dye with lot number 116972 had an expiration date of 03/31/2023. The blue TMD tissue marking dye with lot number 108808 had an expiration date of 10/31/2022. The green TMD tissue marking dye with lot number 81157 had an expiration date of 01/2021. 2. Interview with the lab manager on 09/25/2024 at 2:50 PM confirmed the tissue marking dyes were being utilized for patient testing past their expiration date.</p>