

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0674625	(X3) Date Survey Completed 10/09/2024
Name of Provider or Supplier Integrated Dermatology Of Bountiful	Street Address, City, State 1222 West Legacy Crossing Boulevard, Suite 200, Centerville, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with Laboratory Director (LD), the laboratory failed to ensure that the Yellow Tissue Marking Dye was not used past the expiration date. The laboratory performs approximately 300 histopathology tests annually. Findings Include: 1. Direct observation of tissue marking dyes revealed the Yellow Tissue Marking Dye, lot number 22118, had an expiration date of 04/30/2024 and Yellow Tissue Marking Dye, lot number 115860, had an expiration date of 2/28 /2023 at 10:25 AM on 10/9/2024. 2. Interview with the LD at 10:35 AM on 10/9/2024 confirmed the Yellow Tissue Marking Dye was used for patient testing past the expiration date.</p>