

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0674625	(X3) Date Survey Completed 08/20/2026
Name of Provider or Supplier Integrated Dermatology Of Bountiful	Street Address, City, State 1222 West Legacy Crossing Boulevard, Suite 200, Centerville, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on record review and interview with the General Supervisor (GS), the laboratory failed to evaluate, at least twice annually, the accuracy of their test system for histopathology slide examination for TP#1 (Testing Personnel). The laboratory performs approximately 280 histopathology slide examinations annually. Findings include: 1. Record review revealed that the laboratory failed to verify the accuracy, at least twice annually, of their test system for histopathology slide examination for TP#1 since the last survey performed on 10/09/2024. 2. In an interview on 8/20/2026 at approximately 11:20 am, the GS confirmed the laboratory failed to evaluate, at least twice annually, the accuracy of their test system for histopathology slide examination for the years 2025 and 2026.