

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0687836	(X3) Date Survey Completed 08/05/2019
Name of Provider or Supplier Richard White Parkinson Md	Street Address, City, State 5314 N 250 W, Suite 220, Provo, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on mycology proficiency testing records review, lack of documentation and interview with the office manager and laboratory staff, the laboratory failed to document a review of mycology culture proficiency test results for 4 of 5 events reviewed from the 3rd event of 2017 to the 2nd event of 2019. Findings include: 1. Proficiency testing records review failed to include a signature, initials or date that the reports were reviewed for American Proficiency Institute (API) 3rd event of 2017, the 1st and 2nd events of 2018 and the 1st and 2nd events for 2019. 2. In an interview with the office manager and laboratory staff at approximately 4:50 P.M. staff confirmed the proficiency testing reports lacked documentation they were reviewed to identify any problems with testing.</p>
D5609	<p>HISTOPATHOLOGY CFR(s): 493.1273(e)(f)</p> <p>(e) The laboratory must use acceptable terminology of a recognized system of disease nomenclature in reporting results. (f) The laboratory must document all control procedures performed, as specified in this section.</p> <p>This STANDARD is not met as evidenced by: Based on histopathology staining records review, lack of documentation and confirmation by the office manager, the laboratory failed to document the lot numbers and expiration dates of Hematoxylin (H) and Eosin (E) Stains in use since 08/17/2017.</p>

The laboratory stained approximately 4000 specimens per year Findings include: 1. Histopathology staining records review failed to include the lot number and expiration dates of H & E stains the laboratory used to stain histopathology specimens after the date of 08/17/2017. 2. In an interview conducted on 08/05/2019 at approximately 4:45 P. M., the office manager confirmed the histopathology staining records failed to include the lot number and expiration dates of H & E stains.