

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  46D0687836	<b>(X3) Date Survey Completed</b>  11/13/2025
<b>Name of Provider or Supplier</b>  Richard White Parkinson Md	<b>Street Address, City, State</b>  5314 N 250 W, Suite 220, Provo, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with the Medical Assistant (MA), the laboratory failed to ensure that the Dermatophyte Test Medium (DTM) was not used past the expiration date. The laboratory performs approximately 63 DTM tests annually. Findings Include: 1. Direct observation revealed the DTM, lot number 658018, had an expiration date of 11/2/2025. 2. Interview with the MA at 11:55 AM on 11/13/2025 confirmed the DTM was used for patient testing past the expiration date.</p>