

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0694743	(X3) Date Survey Completed 04/26/2018
Name of Provider or Supplier Joseph M Johnson	Street Address, City, State 1675 N 200 W Ste 9c, Provo, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: The laboratory failed to successfully participate in proficiency testing (PT) for the subspecialty of Bacteriology by failing to achieve a satisfactory score of at least 80% in 2 of 3 consecutive testing events. (See D2028)</p>
D2028	<p>BACTERIOLOGY CFR(s): 493.823(e)</p> <p>Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is</p>

unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on Federal Casper Report 155 and American Proficiency Institute (API) proficiency testing (PT) report review, the laboratory failed to achieve a satisfactory score of at least 80% in 2 of 3 consecutive testing events (event 2 of 2017 and event 1 of 2018) for the sub-specialty of Bacteriology resulting in unsuccessful PT performance. Findings include: 1. Casper and API PT reports document an unsatisfactory score of 45% in Bacteriology for event 2 of 2017. 2. Casper and API PT reports document an unsatisfactory score of 64% in Bacteriology for event 1 of 2018.