

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  46D0694743	<b>(X3) Date Survey Completed</b>  08/24/2023
<b>Name of Provider or Supplier</b>  Joseph M Johnson	<b>Street Address, City, State</b>  1675 N 200 W Ste 9c, Provo, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a routine desk review of the CMS-155 Individual Laboratory Profile report for Proficiency Testing (PT) performance and email communications with the laboratory director, the laboratory failed to achieve satisfactory performance scores for the College of American Pathologists (CAP) PT for Bacteriology testing for two out of three PT events in 2022 and 2023 (event 3 in 2022 and event 1 in 2023). See D2028.</p>
<b>D2028</b>	<p>BACTERIOLOGY CFR(s): 493.823(e)</p>

Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a review of proficiency testing scores from the College of American Pathologists (CAP) and email communications with the laboratory director, the laboratory failed to achieve a score of at least 80% for CAP Proficiency Testing for Bacteriology for event 3 in 2022 and for event 1 in 2023. Findings include: 1. A review of the CMS-155 Individual Laboratory Profile on 8/25/2023, at 11:00 AM, revealed the CAP PT for Bacteriology testing scores for 2022 event 3 was 42%, and for 2023 event 1 was 75%. 2. Email received from the laboratory director on 8/26/2023, at 2:40 PM, confirmed the two consecutive unsuccessful proficiency testing scores for Bacteriology.