

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0694743	(X3) Date Survey Completed 08/26/2024
Name of Provider or Supplier Joseph M Johnson	Street Address, City, State 1675 N 200 W Ste 9c, Provo, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with Laboratory Director (LD), the laboratory failed to ensure that the Stabilized Isotonic Saline was not used past the expiration date. The laboratory performs approximately 800 bacteriology tests annually. Findings Include: 1. Direct observation of Stabilized Isotonic Saline with lot number 2301115 on 8/26/2024 at approximately 1:00 PM , revealed that the Stabilized Isotonic Saline had an expiration date of 7/11/2024. 2. Interview with the LD on 8/26/2024 at approximately 2:00 PM on 8/26/2024, confirmed the Stabilized Isotonic Saline was used for patient testing past the expiration date.</p>