

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0706479	(X3) Date Survey Completed 07/02/2026
Name of Provider or Supplier Dmb Medical	Street Address, City, State 1385 E 750 N, Orem, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on direct observation and interview with the office manger, the laboratory failed to ensure that the Tissue Marking Dyes were not used past the expiration dates. The laboratory performed approximately 3,278 histopathology tests annually. Findings Include: 1. Direct observation at 12:55 PM on 7/2/2026 of Avantik tissue marking dyes revealed 6 tissue marking dyes past the expiration dates. 1. Black Tissue Marking Dye, lot number 176531, had an expiration date of 2025-08-31. 2. Blue Tissue Marking Dye, lot number 182685, had an expiration date of 2025-09-30. 3. Red Tissue Marking Dye, lot number 177813, had an expiration date of 2025-08-31 4. Violet Tissue Marking Dye, lot number 176962 had an expiration of 2025-08-31. 5. Orange Tissue Marking Dye, lot number 167159, had an expiration date of 2025-02-28. 6. Yellow Tissue Marking Dye, lot number 180591, had an expiration date of 2025-09-30. 7. Interview with the office manager at 12:55 PM on 7/2/2026 confirmed the Tissue Marking Dyes were used for patient testing past the expiration dates.