

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  46D0936653	<b>(X3) Date Survey Completed</b>  08/25/2022
<b>Name of Provider or Supplier</b>  Ivan D Flint Md Pc	<b>Street Address, City, State</b>  1250 East 3900 South #450, Salt Lake City, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation of reagents in use and interview with laboratory personnel, the laboratory failed to ensure that expired reagents were removed from use. Findings include: 1. Observation showed on 08/25/2022 at approximately 10:00 AM, one of six tissue marking dyes "purple" had an expiration date of 07/31/2022. 2. In an interview on 08/25/2022 at 10:00 AM, staff confirmed the "purple" dye was expired. Staff confirmed that one of four cases used the purple tissue marking dye while it was expired.</p>