

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D0959710	(X3) Date Survey Completed 12/13/2023
Name of Provider or Supplier Intermountain Life Flight	Street Address, City, State 100 North Mario Capecchi Drive, Salt Lake City, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with technical consultant (TC), the laboratory failed to ensure the Radiometer Safe Clinitubes used for whole blood samples were not used past the expiration date. The laboratory performs approximately 12,909 tests per year. Findings include: 1. Direct observation of inventory on 12/13/2023 at 1:25 P.M. revealed Radiometer Safe Clinitubes with an expiration date of 9/30/2023. 2. Interview with TC on 12/13/2023 at 1:25 P.M. confirmed the Radiometer Safe Clinitubes used for whole blood samples were expired and the laboratory used the tubes to collect patient specimens.</p>