

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  46D0969151	<b>(X3) Date Survey Completed</b>  02/06/2023
<b>Name of Provider or Supplier</b>  Wasatch Pediatrics - Draper Office	<b>Street Address, City, State</b>  114 E 12450 S Ste 100, Draper, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6053</b>	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on a review of competency evaluation documentation and an interview with the Office Manager, the laboratory failed to perform and document semi-annual competency evaluations during 2021-2022 for 5 of 5 Testing Personnel (TP) during their first year of performing moderately complex patient testing Piccolo chemistry analyzer. Findings include: 1. A review of competency evaluation records at approximately 12:00 PM on 2/6/23 revealed that 5 of 5 TP reviewed for 2021-2022, did not have documentation of a semi-annual competency assessment during their first year of performing moderately complex patient testing on the Piccolo chemistry analyzer. 2. An interview with the Office Manager at approximately 12:15 PM on 2/6/2 confirmed that 5 of 5 TP reviewed for 2021-2022, did not have documentation of a semi-annual competency assessment during their first year of performing moderately complex patient testing on the Piccolo chemistry analyzer.</p>