

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D1034137	(X3) Date Survey Completed 01/10/2019
Name of Provider or Supplier Dixie Dermatology	Street Address, City, State 169 W 2710 S Circle Ste 101, St George, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on quality assessment records review, lack of documentation and interview with staff, the laboratory failed to verify potassium hydroxide (KOH) testing at least twice annually in 2017 and 2018. The laboratory tested approximately 50 tests per year. Findings include: 1. Quality assessment records reviewed failed to include documentation the laboratory verified KOH test accuracy at least twice annually in 2017 or 2018. 2. In an interview on 01/10/2019 at approximately 8:30 A.M. the laboratory manager stated they did not record twice annual KOH test test accuracy verification.</p>