

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D1053125	(X3) Date Survey Completed 01/27/2023
Name of Provider or Supplier Intermountain Cedar City Instacare	Street Address, City, State 962 Sage Dr, Cedar City, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with Technical Consultant (TC), the laboratory failed to ensure that Sysmex XN-L Check hematology controls were not used past the expiration date. Findings include: 1. Direct observation of laboratory refrigerator revealed Sysmex XN-L Check hematology controls lot numbers 23231401, 23231402, and 2323103 with an expiration date of January 24, 2023 2. In an interview conducted on January 27, 2023, at approximately 12:30 PM, the TC confirmed the controls were expired and that the laboratory used these controls past their expiration date.</p>