

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  46D1083217	<b>(X3) Date Survey Completed</b>  01/18/2018
<b>Name of Provider or Supplier</b>  Ogden Clinic Utah Hematology Oncology Ogden	<b>Street Address, City, State</b>  5290 South 400 East, Ogden, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2006</b>	<p><b>TESTING OF PROFICIENCY TESTING SAMPLES</b> CFR(s): 493.801(b)</p> <p>The laboratory must examine or test, as applicable, the proficiency testing samples it receives from the proficiency testing program in the same manner as it tests patient specimens. This testing must be conducted in conformance with paragraph (b)(4) of this section. If the laboratory's patient specimen testing procedures would normally require reflex, distributive, or confirmatory testing at another laboratory, the laboratory should test the proficiency testing sample as it would a patient specimen up until the point it would refer a patient specimen to a second laboratory for any form of further testing.</p> <p>This STANDARD is not met as evidenced by: Based on proficiency testing records review, patient test records review, and confirmation by staff, the laboratory failed to ensure proficiency testing was performed by the same testing staff as routinely performed patient testing for 6 of 6 proficiency testing events reviewed from January 2016 to January 2018. Findings include: 1. Proficiency testing records reviewed included documentation the laboratory manager performed complete blood count (CBC) testing for American Proficiency Institute (API) CBC proficiency testing events from January 2016 to January 2018. 2. Patient test records review included documentation at least 11 different testing personnel performed CBC testing. 3.. In an interview with staff on 01/18/2018 staff confirmed the lab manager had performed all 6 of the proficiency testing events over the past 2 years of testing.</p>