

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2082640	(X3) Date Survey Completed 10/10/2018
Name of Provider or Supplier Aspen Dermatology	Street Address, City, State 945 W Hospital Drive, Suite #4, Price, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on reagent records review, lack of documentation, and interview with staff, the laboratory failed to record the lot number and expiration date for Eosin Y stain reagent in use between 03/11/2016 and 09/04/2018. Findings include: 1. The reagent records log failed to include the date, lot number and expiration date of the reagent in use after the expiration date of 03/11/2018 until new reagent lot number was recorded on 09/04/2018 . 2. In an interview with staff on 10/10/2018 at approximately 1:45 P. M. staff stated the laboratory failed to record the reagent that replaced the expired Eosin Y stain prior to the expiration date of 03/11/2018.</p>