

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2086447	(X3) Date Survey Completed 08/19/2025
Name of Provider or Supplier Physician Laboratory Services	Street Address, City, State 2051 E Red Hills Parkway, Suite 7, St George, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: _ Based on a proficiency testing (PT) desk review, the laboratory failed to successfully participate in a proficiency testing program approved by the Department of Health and Human Services (HHS). The laboratory failed to attain a satisfactory performance for the analyte Albumin in two out of three testing events, resulting in an unsuccessful participation. Refer to D2096 _</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p>

(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

_ Based on a desk review of the College of American Pathologists (CAP) proficiency testing (PT) results, the Certification and Survey Provider Enhanced Reporting (CASPER) Report 0155D, and email communications with the technical supervisor (TS), the laboratory failed to achieve satisfactory performance for the analyte Albumin in two consecutive testing events in 2025. _ The findings include: 1. CAP 1st event 2025. a) Albumin = 60% 2. CAP 2nd event 2025. a) Albumin = 20% 3. Casper Report 0155D 1st event 2025. a) Albumin = 60% 4. Casper Report 0155D 2nd event 2025. a) Albumin = 20% 5. The TS confirmed these findings during email communications on August 19, 2025. 6. The laboratory performs approximately 331,784 chemistry tests annually.