

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2100807	(X3) Date Survey Completed 02/23/2023
Name of Provider or Supplier Bahr Dermatology Pc	Street Address, City, State 25 West 500 South, Bountiful, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Laboratory Director, the laboratory failed to establish and follow written policies and procedures to assess annual competency testing for one of one testing personnel performing potassium hydroxide (KOH) slide preparations. The laboratory performs 50 KOH testing per year. Findings include: 1. Record review of personnel documents revealed laboratory failed to document annual competency testing for one of one testing personnel for KOH slide preparations. 2. An interview with the Laboratory Director on February 23 rd, 2023, at approximately 1:30 confirmed failure to document competency testing for one of one testing personnel performing KOH testing.</p>