

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2116123	(X3) Date Survey Completed 02/19/2025
Name of Provider or Supplier Dr Anderson - Foot And Ankle Specialists Of Ut	Street Address, City, State 596 W 750 S, Suite 105, Bountiful, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on lack of documentation for Proficiency Testing (PT), and interview with Technical Supervisor (TS), the laboratory failed to enroll in Proficiency Testing for the specialty of microbiology for Nail PCR Panel and Wound PCR Panel tests performed on the Quant Studio 3 instrument. The laboratory performed approximately 2,850 tests annually. Findings include: 1. A laboratory record review revealed lack of Proficiency Testing documentation for Nail and Wound PCR Panel for the specialty of microbiology testing. 2. Interview with TS on 02/19/2025, at 1:17 PM, confirmed the laboratory was not enrolled in PT for the specialty of microbiology testing for Wound and Nail PCR panel.</p>