Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 46D2122865	**(X3) Date Survey Completed** 01/18/2024
Name of Provider or Supplier Intermountain Utah Valley Clinic Dermatology	**Street Address, City, State** 395 Cougar Blvd 6th Floor, Provo, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on direct observation and interview with Technical Consultant (TC), the laboratory failed to ensure that the Black Tissue Marking Dye was not used past the expiration date. The laboratory performs approximately 850 histopathology tests annually. Findings Include: 1. Direct observation of tissue marking dyes revealed the Black Tissue Marking Dye, lot number 048108, had an expiration date of 07/2018 at 12:10 PM on 1/18/2024. 2. Interview with the TC at 12:10 PM on 1/18/2024 confirmed the Black Tissue Marking Dye was used for patient testing past the expiration date.