

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  46D2139397	<b>(X3) Date Survey Completed</b>  01/13/2021
<b>Name of Provider or Supplier</b>  Cvmc Santaquin Medical Clinic	<b>Street Address, City, State</b>  210 E Main Street, Santaquin, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and confirmation by staff, the laboratory failed to ensure the heavy metal specimen collection evacuated tube was not used past the expiration date. The laboratory performed approximately 2 specimen collections per year using the dark blue specimen collection tubes. Findings include: 1. Direct observation of specimen collection laboratory area included expired dark blue topped evacuated specimen container tubes in 2 specimen collection containers. The tube labeled the expiration date as 12/31/2020. The date of survey was 01/13/2021. 2. In an interview conducted on 01/13/2021 at approximately 3:00 P.M. staff confirmed the collection tubes were expired and in date replacement tubes were not on the premises.</p>